Standard Form Number: SF-GOOD-04

Revised on: May 24, 2004

## Republic of the Philippines

Jose R. Reyes Memorial Medical Center (02) 711-9491 loc. 245 / 8562-5338

# MINUTES OF THE MEETING PRE-BIDDING CONFERENCE OF VARIOUS MEDICAL EQUIPMENT 09 June 2022

### **Present Were:**

## HBAC-A2 Chairperson:

1. Dr. Emelito O. Valdez - Tan

## HBAC-A2 Vice-Chairperson:

2. Ms. Natividad Sarmiento

#### **HBAC-A2 Members:**

- 3. Dr. Juanita Castillo Cariño
- 4. Ms. Joseph T. Gatchalian
- 5. Mr. Arlie Felvin Gacias

## **Technical Working Group**

- 1. 10-Header Microscope Dr. Mary Darice Wong, Dr. Joseph Tayag & Dr. Elisa Rae Coo
- 2. Various Pedia Equipment (For Milk Bank) Dr. Donna Cagalawan & Dr. Jeanette Turing
- 3. Operating Table Emelita F. Chang
- 4. Autoclave Ms. Cecilia F. Castro
- 5. Mobile Digital X-ray Dr. Neil Martin Sese & Mr. Rey Anthony Ababon

## **AHW Representative**

1. Mr. Arnold Pascual

## Bidder/s:

- 1. Mr. Edgar Liu Prime Medix Enterprises
- 2. Ms. Jocelyn V. Gomesera, Nazaat Jailani, Sahar L. Gregorio, Maryjo Tamayo Zafire Distributor
- 3. David Tan Respicare Sales
- 4. Jeffrey Flores JQM Enterprises Corporation
- 5. Ella Uy Assurance Controls
- 6. Ma. Theresa Paragile Med Care Supplies
- 7. Patrick Taculod Guill Bern Corp.
- 8. Eric Martinez, Myra Garcia Omnibus
- 9. Mary Rose Angeles Babymama, Inc.
- 10. Justin Clemena MMJ Biosystem
- 11. Arnold B. Mendoza Shimadzu Phils.
- 12. Joselito Lopez Instrumix Supplier
- 13. Eric Aringo AMI Equipment
- 14. Robert Ablog Goldquest
- 15. Rose Suarez, Dondaryll Goopio Patient Care
- 16. Maynard Heneria Variance
- 17. Jhon Ritz Valdez Bullseye Solutions

The meeting of the Bids and Awards Committee (BAC), held at Derma Conference Room was called to order at 2:15pm, of 09 June 2022 and was presided by the BAC Chairman, Dr. Emelito O. Valdez - Tan.



Page \_\_\_\_ of \_\_\_\_\_\_

Standard Form Number: SF-GOOD-04 Revised on: May 24, 2004

## I. Business Matter

a) The Schedule of Requirements, Approved Budget for the Contract (ABC) and Delivery days was flashed in the screen and reviewed by the Bidder/s.

# 10-HEADER MICROSCOPE WITH SLIDE SHARER AND DIGITAL STORAGE

- ➤ The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.
  - a High luminosity LED system with excellent color rendering performance that provides brightness and color reproduction <u>equal to or better than that of a</u> 100 W halogen lamp. "equal to or better than that of a" change to <u>"at least 100 W"</u>.
  - For the high luminosity LED "or live streaming features" was added.
  - For specimen holder for BX stage, "BX" was deleted.
  - For all software updates are included, "(slide share)" was added.
  - For software support, "(slide share)" was added.
  - Clarification for Years of Hosting Subscription was set to <u>"for 3 Years"</u> temporarily as per Dr. Wong but still for approval of their Department Chairman.
  - Guaranteed server uptime was change to "99.95%".
  - In User's Manual in English Language, "/ Operator's Manual" was added.
  - In Service Manual, "/ Operator's Manual" was added.

### MOBILE DIGITAL X RAY

- ➤ The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.
  - At least "53-195" SID to floor. Unit was for clarification.
  - At least 1300mm Tube arm reach total, was change to <u>"1280mm"</u>
  - Indication of lights (sides, column top), "or its equivalent" was added
  - Size of "Movable standard lead barrier with lead glass" was for clarification.
  - Years of Warranty was for clarification if "<u>can be 2 years on parts and service and additional year for battery"</u>.

### **AUTOCLAVE STERILIZER WITH REVERSE OSMOSIS SYSTEM**

- ➤ The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.



# OPERATING TABLE, HEAVY DUTY, MANUAL, ELECTRO-HYDRAULIC

- > The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - In Height adjustment: 20 inches 44 inches, "<u>lowest adjustment of</u>" was added.
  - In Dimension: at least 200 cm x 55 cm (LxW) with thickness of at least three (3) inches <u>"(7.62cm)"</u> was added and <u>"at least"</u> before 55 cm was deleted.
  - In One (1) pair arm boards with three (3) inches mattress, "[7.62cm]" was added
  - In One (1) piece head board with THREE (3) INCHES, "(7.62cm)" was added
  - In One (1) pair leg board with THREE (3) INCHES, "[7.62cm]" was added

# **PASTEURIZER**

- ➤ The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.
  - In Rapid Cooling, "Yes" was deleted.
  - In Foil Seal Bottle: at least 200 pieces, "Option: Optional," was deleted.

## MEDICAL GRADE REFRIGERATOR (UPRIGHT)

- ➤ The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.
  - In Construction: Stainless steel, "inner lining" was added.
  - "R290, R407D" was for clarification if can be added to Refrigerant: R600a, CFC- and HCFC-free.
  - Temperature Range was also for clarification if can be adjusted to "+4°C to +15°C".
  - For Interior Light: LED with separate switch, "or its equivalent" was added.
  - For Self-Closing Door: with fitted lock, "Hinge Type" was added.
  - In Door Lock, "Optional" was deleted.

# MEDICAL GRADE FREEZER (UPRIGHT)

- ➤ The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.
  - Construction: Stainless steel was for clarification if it can be Electrogalvanized Steel.
  - In Power Requirement: 220v-240v, "60Hz" was added
  - For Interior Light: LED with separate switch, "or its equivalent" was added.



Standard Form Number: SF-GOOD-04

Revised on: May 24, 2004

- In Door Lock, "Optional" was deleted.

## **LAMINAR FLOW CABINET**

- The following specifications were discussed and clarified by the Bidder/s with BAC & TWG End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
  - Delivery days was requested by the bidder to change to <u>"45 Calendar Days"</u> to which the BAC and End-user agreed.
  - For Internal Work Area width x depth x height was clarified.
  - In Display, "LCD" was added.
  - For Front Window, tempered glass was for clarification if can be polycarbonate.

There, having no other remaining topics for discussion, the meeting adjourned at 5:25pm.

Prepared by:

ADRIELE F. DASMARIÑAS

BAC Secretariat

EMELITO VALDEZ - TAN, MD

Chairperson – HBAC – A2

Noted b

CHEL/may

Page f of 4