


**JOSE R. REYES MEMORIAL MEDICAL CENTER**

Rizal Avenue, Sta. Cruz, Manila

**BAC SECRETARIAT****SEALED BID**SEALED BID NO.: HBAC-A2-2020-06-0701Date: 01 June 2020**Gentlemen:**

Please quote your price for item/s indicated below. Price quoted is final, and without condition and inclusive of all taxes. Bid Proposal should be submitted in sealed envelope, properly labelled, indicating the date and Sealed Bid No., not later than 2:00pm of 04 June 2020 at the BAC Secretariat Office and thereafter publicly opened. Any change/erasure must be duly initialed otherwise, the bid shall be disqualified. Attendance of authorized representative is not mandatory on bid opening. In accordance with Section 53.2 of the Revised Implementing Rules and Regulation of the Republic Act 9184.

THE COMMITTEE RESERVES THE RIGHT TO REJECT ANY AND ALL BIDS, DECLARE A FAILURE OF BIDDING AT ANY TIME PRIOR TO THE CONTRACT AWARD, OR NOT TO AWARD THE CONTRACT, WITHOUT THEREBY INCURRING ANY LIABILITY, AND MAKE NO ASSURANCE THAT A CONTRACT SHALL BE ENTERED INTO AS A RESULT OF THE BIDDING.

For:   
**E MELITO O. VALDEZ-TAN, MD**  
Chairman, HBAC-A2

PR No.: 2020-05-2924 (HEMC)

**PLEASE INDICATE BRAND / MODEL IF APPLICABLE**

Qty.	Unit	Description/Particulars	ABC	Unit Price	Total Price
1	Unit	<b>2D ECHOCARDIOLOGY SYSTEM (2D ECHO)</b> <b>Branded, Brand New</b> <b>Physical Characteristics</b> <ul style="list-style-type: none"><li>At least 15-inch monitor that can be tilted up to -10 degrees and +65 degrees and can be rotated by at least -90 and +90 degrees</li><li>With keyboard for easy patient data annotation and report entries.</li><li>With trackball for fast and easy control of equipment functions</li><li>With built - in image printer and built - in digital image archiving</li><li>With built - in DVD writer and with USB port</li><li>Mounted on a cart with 4 anti-static wheels with brakes</li></ul> <b>Power Supply</b> <ul style="list-style-type: none"><li>220V, 60Hz with atleast 2KVA Uninterruptible Power Supply (UPS) that can provide back-up power of up to 30 minutes</li></ul> <b>Technical Characteristics:</b> <ul style="list-style-type: none"><li><b>Displays / Indicators:</b><ul style="list-style-type: none"><li>➤ <b>Display Channel</b> Electrocardiogram (ECG) amplifier and display in addition to at least one other physiological channel amplifier display</li><li>➤ <b>Adequate Resolution</b><ul style="list-style-type: none"><li>✚ <b>Adult</b><ul style="list-style-type: none"><li>○ Axial less than 0.5mm at all depths and less than 2mm in focal zones</li><li>○ Slice thickness of less than 8mm at all depths</li></ul></li><li>✚ <b>Pediatric</b><ul style="list-style-type: none"><li>○ Axial less than 0.3mm lateral less than 3mm at all depths and less than 1mm in focal zone</li><li>○ Slice thickness of less than 5mm at</li></ul></li></ul></li></ul></li></ul>	2,600,000.00		





# JOSE R. REYES MEMORIAL MEDICAL CENTER

Rizal Avenue, Sta. Cruz, Manila

## BAC SECRETARIAT

### SEALED BID

		<p>all relevant depths</p> <ul style="list-style-type: none"><li>✦ <b>Multiple Image Display</b><ul style="list-style-type: none"><li>○ Capable of displaying at least two images in same mode simultaneously</li></ul></li><li>✦ <b>Multi - mode Display</b><ul style="list-style-type: none"><li>○ Simultaneous display of B, M spectral, CD and power Doppler modes</li></ul></li><li>• <b>Functions</b><ul style="list-style-type: none"><li>➤ Brightness (B) mode with tissue harmonic imaging</li><li>➤ Focal level / Focal depth adjustment</li><li>➤ Adjustment of frequency range on all transducers</li><li>➤ Capability to select at least three transducers without physical removal and reconnection</li><li>➤ Gain and Time Gain Compensation (TGC) control</li><li>➤ Operator-controlled multiple and adjustable focal zones</li><li>➤ Provision for scanning pre-sets</li><li>➤ Measurement of linear and curved distances, areas and volumes</li><li>➤ Look-up tables to link measurements to relevant clinical applications</li><li>➤ Cine-loop capability</li><li>➤ Automatic and manual calculation of waveform indices</li><li>➤ Capability of operator application presets</li><li>➤ Magnification using both read and write zoom</li><li>➤ Freeze and replay capabilities</li><li>➤ Capability of displaying at least 5 measurements simultaneously</li><li>➤ Patient identification and entry of other relevant clinical information</li><li>➤ Capability of displaying, Patient name, Institution Name, Transducer type, date, time, frame rate / persistence, Thermal Index (TI) and Mechanical Index (MI)</li></ul></li><li>• <b>Modes</b><ul style="list-style-type: none"><li>➤ 2 Dimension (2D), Motion (M) mode</li><li>➤ Anatomical M mode</li><li>➤ Continuous Wave (CW) Doppler</li><li>➤ Pulsed - Wave (PW) Doppler</li><li>➤ Color Doppler</li></ul></li><li>• <b>Spectral Doppler:</b> Range gate accuracy of less than 1mm</li><li>• <b>Color Doppler:</b> Adjustable Thump Filter</li><li>• <b>Microbubble Imaging:</b> Suitable scanning mode capability</li><li>• <b>Adequate Penetration:</b> At least 15cm of normal tissue</li><li>• <b>Transducers</b> (the transducers must have beam steering capability)<ul style="list-style-type: none"><li>➤ Adult Echo Transducer</li><li>➤ Pediatric Echo Transducer</li><li>➤ Curved array probe for fetal echocardiography</li><li>➤ Adult trans-esophageal probe</li></ul></li></ul>		
--	--	--	--	--





# JOSE R. REYES MEMORIAL MEDICAL CENTER

Rizal Avenue, Sta. Cruz, Manila

## BAC SECRETARIAT

### SEALED BID

➤ Neonatal transthoracic probe

**Note: The bidder must declare all transducer frequency ranges and applications in the bid specifications**


- Software and applications for trans-esophageal and echocardiography (including Doppler) and other cardiac function analysis applications (such as lumen border measurement and analysis, mitral valve analysis, etc.) must be installed.
- The equipment must be compliant with Digital Imaging and Communications in Medicine (DICOM) and must be compatible with color Inkjet printers
- **Accessories**
  - Thermal Printer
  - DVD Burner

- Non - removable embossed DOH letters on the visible part of the equipment

**Documentary Requirements:**

- Mayor's Permit
- Omnibus Sworn Statement
- Income / Business Tax Return
- Product brochure or technical data sheets of the equipment showing the technical specifications in English Language
- Valid and current Certificate of Compliance of Manufacturer of the Equipment with the latest version of ISO 13485: Quality Management System - Requirements for Regulatory Purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body / Agency
- Valid Marketing Authorization, Registration Approval or Free Sale Certificate for the equipment issued by the Health Authority in the Country of Origin.
- Valid Certificate of Distributorship (as first Tier Distributor) issued by the Manufacturer of the Equipment Authorizing the bidder to sell / distribute the offered equipment
- List and address of the equipment manufacturer's branch office, sales office and / or distributor's office in any of the following:
  - Western Europe;
  - USA or Canada and;
  - Japan
- Proof (such as sales invoice) that the Brand of the equipment has been sold to other health facilities in the Philippines.
- Notarized Certificate from the Bidder:
  - That the brand of the equipment has been in the local and / or international market for at least ten (10) years.
  - That the equipment and its accessories are brand new, unused, not-discontinued models and were not subjected to any product recall
  - Bidder's valid and current License to





# JOSE R. REYES MEMORIAL MEDICAL CENTER

Rizal Avenue, Sta. Cruz, Manila

## BAC SECRETARIAT

### SEALED BID

Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted:

- ✚ Copy of expired LTO
- ✚ Application for renewal
- ✚ Official Receipt as proof of payment for the renewal of LTO

**Requirements if Awarded the Contract**  
*(Please attach a Certification that the following will be provided / will be followed upon delivery of the item)*

- **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within Seven (7) Calendar Days upon receipt of the Notice to Proceed.
- **Testing:** The equipment and accessories must be functioning with no physical damage and / or defect. A Performance Evaluation must be conducted by the Center for Device Regulation Radiation Health and Research, Food and Drug Administration (FDA) or its authorized representative. The bidder shall be the one to apply for the Performance Evaluation in FDA. Application fees and other expenses for the conduct of the Performance Evaluation shall be borne by the bidder.
- **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff. The training must consist of familiarization of the equipment controls, displays, functions, settings, etc. Specialized applications training for Doppler and 2D echo and other cardiac capabilities of the ultrasound machine.
- **Warranty:** Warranty certificate for three (3) years on Parts and on Services. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The Warranty period shall commence from the date of acceptance by the end-user after testing and commissioning
- Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of equipment breakdown from the end-user. The undertaking shall include a

8 Jmas





# JOSE R. REYES MEMORIAL MEDICAL CENTER

Rizal Avenue, Sta. Cruz, Manila

## BAC SECRETARIAT

### SEALED BID

		<p>statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <ul style="list-style-type: none"><li>• <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:<ul style="list-style-type: none"><li>➤ Service Manual in English Language</li><li>➤ Operations manual in English Language</li></ul></li></ul>			
--	--	---	--	--	--

#### Check Mode of Payment:

\_\_\_ 1. Terms/Deferred Payment

\_\_\_\_\_  
Name of Company

\_\_\_\_\_  
Address/TIN No.

**No Clearing time before delivery, No pick-up**

\_\_\_\_\_  
Authorized signature over printed name/ Telephone no.

**NOTE: Signing for these request for quotation, is conforming to our payment terms you check.**

**MAXIMUM DELIVERY PERIOD; WITHIN 7 WORKING DAYS UPON RECEIPT OF NTP**

**HOSPITAL BIDS AND AWARDS COMMITTEE FOR SUPPLIES, MATERIALS & EQUIPMENT**

Chairman

Vice-Chairman

Member

Member

Member

rich/mau